

THE FOLLOWING PAGE CONTAINS THE CONNECTION DETAILS

It is your responsibility to ensure that the transducer identification matches the connection details, including any RDP modification (T.M) references.

Please ensure that the transducer has not been modified or fitted with a cable or extension.

Particularly on DC-DC powered transducers, incorrect connection may cause **IRREPERABLE DAMAGE** to the transducer which is **NOT COVERED BY WARRANTY**


**IF YOU ARE IN ANY DOUBT PLEASE CONTACT US**



# ACW SERIES

(LVDT - OPTIONS 2,3,4,5,8,10,11)

## CONNECTION DETAILS



**ENERGISING SUPPLY :** 0.5 to 7V r.m.s. 5kHz optimum.  
**SENSITIVITY :** Refer to RDP Calibration Certificate.  
**CALIBRATED :** 5V 5kHz energisation into 100k Ohms.  
**PHASING :** YELLOW and GREEN are in phase with the armature fully extended.

**RDP U.K.**  
**Tel 01902 457512**  
**Fax 01902 452000**



Certificate No. FM 13141  
BS EN ISO 9001:2000

**MADE IN THE  
UNITED KINGDOM**

Doc. Ref. WDUK1401  
ISSUE 4 - 16 AUG 2004

**RDP U.S.A.**  
**Tel 610 469 0850**  
**Fax 610 469 0852**

## Distribuidor

Brasil e América do Sul

### CONTATO

Endereço

Rua Sete de Setembro, 2656 - Centro  
13560-181 - São Carlos - SP - Brasil

Telefone

+ 55 (16) 3371-0112  
+ 55 (16) 3372-7800

Internet

[www.metrolog.net](http://www.metrolog.net)  
[metrolog@metrolog.net](mailto:metrolog@metrolog.net)

