


Calibration Sheet

Model: LRI-19-100-R-00-20-A
Serial Number: 2417-0652
Date: 10/27/2017
Linearity: +/- 0.095 %

Length (in)	Current (mA)
-0.000	4.0399
0.400	5.6278
0.800	7.2181
1.200	8.8073
1.600	10.4216
2.000	12.0313
2.400	13.6163
2.800	15.2256
3.200	16.8243
3.600	18.4268
4.000	20.0136



SAMPLE COPY ONLY

DISTRIBUTOR / DISTRIBUIDOR

Brazil and South America / Brasil e América do Sul



Address / Endereço:

Rua Sete de Setembro, 2656
13560-181 - São Carlos - SP
Brazil / Brasil

Phone / Telefone:

+55 (16) 3371-0112
+55 (16) 3372-7800

Internet:

www.metrolog.net
metrolog@metrolog.net

ALLIANCE SENSORS GROUP
A DIVISION OF H.G. SCHAEVITZ LLC

Alliance Sensors Group
102 Commerce Drive, Unit 8
Moorestown, New Jersey 08057 USA
Ph: 856-727-0250
www.alliancesensors.com
info@alliancesensors.com