

THE FOLLOWING PAGE CONTAINS THE CONNECTION DETAILS

It is your responsibility to ensure that the transducer identification matches the connection details, including any RDP modification (T.M) references.

Please ensure that the transducer has not been modified or fitted with a cable or extension.

Particularly on DC-DC powered transducers, incorrect connection may cause **IRREPERABLE DAMAGE** to the transducer which is **NOT COVERED BY WARRANTY**


IF YOU ARE IN ANY DOUBT PLEASE CONTACT US

DESPATCH:- INCLUDE CD1013 IF TRANSDUCER HAS NO CABLE FITTED.



SSA SERIES (LVDT)

CONNECTION DETAILS



ENERGISING SUPPLY : 0.5 to 7V r.m.s. 5kHz optimum.
SENSITIVITY : Refer to RDP Calibration Certificate.
LOAD IMPEDANCE : 100k Ohms optimum
PHASING : WHITE and GREEN are in phase with the armature fully extended.

For transducers without cable see CD1013 for fitting instructions

RDP U.K.
Tel 01902 457512
Fax 01902 452000



Certificate No. FM 13141
BS EN ISO 9001:2000

**MADE IN THE
UNITED KINGDOM**

Doc. Ref. WDUK1350
ISSUE 3 - 14 SEP 2004

RDP U.S.A.
Tel 610 469 0850
Fax 610 469 0852

Distribuidor

Brasil e América do Sul

CONTATO

Endereço

Rua Sete de Setembro, 2656 - Centro
13560-181 - São Carlos - SP - Brasil

Telefone

+ 55 (16) 3371-0112
+ 55 (16) 3372-7800

Internet

www.metrolog.net
metrolog@metrolog.net

